

SAMPLE NUMBER	REF NO	SUMMARY OF RESULTS IN % CORRECTED FOR BLANK AND BIAS								SET 78	ANALYST & F. RALPH	TOTAL							
		SIO2	AL203	FE203	MGO	CAO	NA20	K20	TIO2				+H2O	-H2O	CO2	P205	S	FEO	FE203
SAMPLE		REF	SI02	AL203	FE203	MGO	CAO	NA20	K20	TIO2	MNO	+H2O	-H2O	CO2	P205	S	FEO	FE203	TOTAL
SAMPLE	20838	1	63.13	16.13	5.95	0.15	0.44	6.223	0.883	0.840	<0.031	0.91	0.31	0.15	0.25	4.76	0.39	5.51	100.12
SAMPLE	20839	2	57.25	23.88	4.35	0.49	0.48	0.566	5.208	1.209	0.039	3.08	0.74	0.45	0.18	1.85	1.39	2.01	99.57
SAMPLE	20840	3	55.30	20.00	7.92	1.53	0.43	1.084	3.235	1.082	0.098	3.53	0.91	0.15	0.22	4.76	1.73	5.99	100.06
SAMPLE	20841	4	65.49	20.10	5.45	0.90	0.22	0.293	2.375	1.232	0.175	3.03	0.33	0.15	0.14	0.04	4.54	0.40	99.42
SAMPLE	20842	5	62.34	20.12	1.88	1.13	1.29	6.684	0.731	1.509	0.082	1.73	0.57	0.29	0.45	1.06	0.56	1.26	99.80
SAMPLE	20843	6	52.40	20.19	9.12	1.65	1.19	1.774	3.498	1.061	0.163	3.68	0.86	0.15	0.93	4.30	3.53	5.20	100.59
SAMPLE	BLANK	7	< 0.42	< 0.04	< 0.05	< 0.02	< 0.02	< 0.007	< 0.012	< 0.013	< 0.031	0.	0.	0.	0.	0.	0.	0.	0.61
SAMPLE	20844	8	63.00	19.25	7.77	1.51	0.42	0.382	1.870	1.032	0.096	3.53	0.43	0.29	0.14	0.17	6.05	1.05	99.23
SAMPLE	20845	9	61.93	19.18	5.42	1.04	0.40	0.149	5.307	1.105	0.091	2.22	0.92	0.15	0.30	0.70	2.96	2.13	98.58
SAMPLE	SY-2	10	59.65	12.10	6.2	2.65	7.99	4.309	4.547	0.151	0.321	0.46	0.23	0.46	0.44	0.01	3.54	2.33	99.19
SAMPLE	20846	11	60.76	19.59	5.10	0.39	0.27	0.227	4.038	1.052	< 0.031	3.17	0.41	0.44	0.11	1.94	2.22	2.63	97.28
SAMPLE	20847	12	64.26	17.33	5.62	0.48	0.24	0.134	5.254	1.067	< 0.031	2.19	0.17	0.15	0.18	4.15	1.09	4.40	101.13
SAMPLE	20848	13	56.16	19.68	8.05	1.20	0.40	0.640	4.245	0.911	0.043	2.99	0.51	0.29	0.16	4.03	3.26	4.44	99.64
SAMPLE	22260	14	45.75	15.24	12.29	6.19	9.65	2.733	0.358	2.081	0.209	3.76	0.	0.33	0.18	0.07	4.20	7.62	100.37
SAMPLE	22261	15	84.88	5.51	1.57	1.00	2.06	0.025	0.791	0.177	0.110	0.94	0.	1.47	0.92	0.01	1.30	0.13	99.32
SAMPLE	22262	16	50.35	13.80	9.27	6.30	9.92	2.635	0.127	1.810	0.155	3.34	0.	1.76	0.32	0.01	7.05	1.43	99.00
SAMPLE	SY-2	17	59.87	12.06	6.25	2.68	7.93	4.351	4.592	0.143	0.325	0.46	0.23	0.46	0.44	0.01	3.54	2.31	99.41
SAMPLE	BLANK	18	< 0.42	< 0.04	< 0.05	< 0.02	< 0.02	< 0.007	< 0.012	< 0.013	< 0.031	0.	0.	0.	0.	0.	0.	0.	0.61

896173

Chu-chua

RELATIVE STANDARD

DEVIATION %

INITIAL FLAME = 3.836 7.561 2.721 3.254 ***** 2.428 1.638 9.879 3.649

FINAL FLAME = 1.575 0.653 1.233 0.681 1.838 1.334 0.804 5.786 *****

ABOUT LINE = 0.697 0.839 0.734 0.404 0.802 0.494 0.349 ***** 1.363

LIMIT DETECTION= 0.421 0.037 0.051 0.017 0.024 0.007 0.012 0.013 0.031

POLYNOMIAL ORDER

DRIFT = 3 7 3 4 3 2 5 4 2

CALIBRATION = 3 3 5 4 3 3 3 2 2

RELATIVE STANDARD DEVIATION ON STANDARD SY-2

BETWEEN SAMPLES=<0.879 <1.586 <2.856 0.742 <4.907 <2.704 <1.817 <***** <1.789

MEASUREMENT

OBTAINED = 0.194 0.350 0.630 0.109 1.082 0.596 0.401 9.613 0.395

EXPECTED = 1.035 0.482 0.825 0.446 1.122 0.923 0.641 5.468 9.257

MEAN ON SY-2 = 59.76 12.08 6.25 2.67 7.96 4.328 4.567 0.151 0.323

ACCEPTED MEAN = 59.76 12.08 6.26 2.68 7.96 4.33 4.49 0.15 0.32

SAMPLE NUMBER	REF NO	SUMMARY OF STANDARD DEVIATION OF RESULTS												TOTAL					
		SIO2	AL203	FE203	MGO	CAO	NA20	K20	TIO2	MNO	+H2O	-H2O	CO2	P205	S	FEO	FE203		
SAMPLE		REF	SI02	AL203	FE203	MGO	CAO	NA20	K20	TIO2	MNO	+H2O	-H2O	CO2	P205	S	FEO	FE203	TOTAL
SAMPLE	20838	1	0.685	0.098	0.059	0.005	0.010	0.077	0.008	0.033	0.014							0.700	
SAMPLE	20839	2	0.619	0.375	0.045	0.005	0.010	0.007	0.042	0.048	0.018							0.728	
SAMPLE	20840	3	0.596	0.150	0.076	0.010	0.010	0.011	0.021	0.043	0.016							0.622	
SAMPLE	20841	4	0.713	0.155	0.055	0.007	0.008	0.005	0.016	0.049	0.020							0.734	
SAMPLE	20842	5	0.673	0.145	0.034	0.008	0.019	0.084	0.007	0.060	0.016							0.697	
SAMPLE	20843	6	0.567	0.151	0.085	0.010	0.018	0.017	0.023	0.042	0.020							0.595	
SAMPLE	BLANK	7	0.206	0.020	0.029	0.006	0.008	0.004	0.007	0.007	0.015							0.210	
SAMPLE	20844	8	0.682	0.138	0.075	0.009	0.010	0.005	0.013	0.041	0.013							0.702	
SAMPLE	20845	9	0.670	0.137	0.054	0.007	0.009	0.004	0.041	0.044	0.016							0.689	
SAMPLE	SY-2	10	0.643	0.066	0.062	0.015	0.097	0.043	0.031	0.009	0.031							0.660	
SAMPLE	20846	11	0.658	0.151	0.051	0													

SUMMARY OF RESULTS IN % CORRECTED FOR BLANK AND BIAS SET 74 ANALYST CHAUDHRY

SAMPLE NUMBER	REF NO	SIO2	AL203	FE203	MGO	CAO	NA20	K20	TIO2	MNO	+H2O	-H2O	CO2	P205	S	FEO	FE203	TOTAL
CC 14-20 -49 74.5	SAMPLE 22369 1	47.58	15.61	11.58	6.38	11.65	1.936	0.158	1.945	0.180	0.	0.	0.	0.	0.	0.	0.	97.03
3CC4-150 100 115	SAMPLE 22370 2	47.28	15.55	11.68	6.67	11.17	2.185	0.079	1.955	0.190	0.	0.	0.	0.	0.	0.	0.	96.76
SAMPLE 22371 3	49.97	14.69	10.57	6.25	10.21	3.445	0.101	1.889	0.182	0.	0.	0.	0.	0.	0.	0.	97.31	
SAMPLE 22374 4	48.18	14.78	10.98	6.54	10.54	3.130	0.053	1.794	0.192	0.	0.	0.	0.	0.	0.	0.	96.40	
SAMPLE 22375 5	50.78	14.27	10.04	6.06	9.06	3.668	0.075	1.782	0.193	0.	0.	0.	0.	0.	0.	0.	95.92	
SAMPLE 22376 6	47.66	15.58	11.51	7.24	10.41	2.638	0.138	1.927	0.198	0.	0.	0.	0.	0.	0.	0.	97.30	
SAMPLE BLANK 7	< 0.13	< 0.06	< 0.03	< 0.02	< 0.02	< 0.030	< 0.013	< 0.054	< 0.007	0.	0.	0.	0.	0.	0.	0.	0.37	
SAMPLE 22378 8	45.15	14.48	10.96	8.24	6.30	1.355	1.083	1.697	0.170	0.	0.	0.	0.	0.	0.	0.	89.44	
SAMPLE 22379 9	58.25	15.38	17.67	5.89	0.91	< 0.030	0.025	0.279	0.044	0.	0.	0.	0.	0.	0.	0.	88.48	
SAMPLE SY-2 10	59.65	12.14	6.23	2.64	7.96	4.259	4.445	0.150	0.326	0.46	0.23	0.46	0.44	0.01	3.54	2.30	99.02	
SAMPLE 22377 11	43.16	13.90	10.23	6.98	8.11	0.110	1.472	1.728	0.173	0.	0.	0.	0.	0.	0.	0.	85.87	
SAMPLE 22378 12	44.54	13.19	10.46	6.43	9.45	1.680	0.652	1.685	0.194	0.	0.	0.	0.	0.	0.	0.	86.29	
SAMPLE 22379 13	45.21	13.49	10.77	7.14	8.33	1.918	0.568	1.675	0.253	0.	0.	0.	0.	0.	0.	0.	89.36	
SAMPLE 22381 14	47.57	15.17	11.23	6.63	11.90	2.539	0.049	1.809	0.198	0.	0.	0.	0.	0.	0.	0.	97.09	
SAMPLE 22382 15	45.36	13.88	10.28	5.69	10.04	2.819	0.047	1.678	0.169	0.	0.	0.	0.	0.	0.	0.	89.95	
SAMPLE 22383 16	49.39	14.60	11.25	6.38	9.50	3.811	0.074	1.797	0.190	0.	0.	0.	0.	0.	0.	0.	96.98	
SAMPLE 22384 17	42.16	14.43	10.49	6.30	7.64	0.348	1.759	1.613	0.141	0.	0.	0.	0.	0.	0.	0.	84.88	
SAMPLE BLANK 18	< 0.13	< 0.05	< 0.03	< 0.02	< 0.02	< 0.030	< 0.013	< 0.054	< 0.007	0.	0.	0.	0.	0.	0.	0.	0.37	
SAMPLE 22385 19	60.51	10.62	3.92	29.08	0.04	0.072	< 0.013	0.065	< 0.007	0.	0.	0.	0.	0.	0.	0.	94.32	
SAMPLE 22386 20	45.09	13.60	10.43	6.33	7.12	0.243	1.252	1.672	0.159	0.	0.	0.	0.	0.	0.	0.	85.90	
SAMPLE 22387 21	49.67	14.17	10.76	5.49	10.69	2.949	0.034	1.641	0.179	0.	0.	0.	0.	0.	0.	0.	95.58	
SAMPLE SY-2 22	59.82	12.22	6.16	2.54	7.98	4.284	4.518	0.153	0.324	0.46	0.23	0.46	0.44	0.01	3.54	2.23	99.36	
SAMPLE 22389 23	46.93	15.56	11.76	7.24	9.68	2.936	0.109	1.952	0.205	0.	0.	0.	0.	0.	0.	0.	96.38	
SAMPLE 22390 24	47.33	15.32	11.94	7.05	7.39	3.302	0.417	1.828	0.212	0.	0.	0.	0.	0.	0.	0.	94.78	
SAMPLE 22391 25	47.02	14.40	10.16	6.62	4.99	0.394	1.664	1.803	0.178	0.	0.	0.	0.	0.	0.	0.	87.23	
SAMPLE 22393 26	42.37	15.02	12.08	6.90	6.93	2.865	0.412	1.876	0.190	0.	0.	0.	0.	0.	0.	0.	88.66	
SAMPLE 22394 27	43.78	13.51	9.27	5.62	7.55	0.203	2.228	1.652	0.155	0.	0.	0.	0.	0.	0.	0.	83.97	
SAMPLE 22395 28	43.10	13.12	10.14	6.39	4.69	0.089	1.081	1.656	0.152	0.	0.	0.	0.	0.	0.	0.	80.51	
SAMPLE 22396 29	47.48	15.98	11.42	6.64	9.72	2.653	0.556	1.926	0.191	0.	0.	0.	0.	0.	0.	0.	96.56	
SAMPLE BLANK 30	< 0.13	< 0.06	< 0.03	< 0.02	< 0.02	< 0.030	< 0.013	< 0.054	< 0.007	0.	0.	0.	0.	0.	0.	0.	0.37	
SAMPLE 22397 31	46.61	15.64	12.19	7.22	9.27	3.361	0.147	1.900	0.206	0.	0.	0.	0.	0.	0.	0.	96.54	
SAMPLE 22398 32	45.87	15.50	11.13	6.70	10.94	2.200	0.423	1.906	0.196	0.	0.	0.	0.	0.	0.	0.	95.86	
SAMPLE 22399 33	47.89	14.95	11.00	6.30	11.91	1.625	0.133	1.893	0.192	0.	0.	0.	0.	0.	0.	0.	95.89	
SAMPLE SY-2 34	60.35	11.87	6.24	2.67	7.91	4.304	4.481	0.170	0.328	0.46	0.23	0.46	0.44	0.01	3.54	2.31	99.53	
SAMPLE 22401 35	71.69	8.91	4.66	2.92	0.59	< 0.030	0.056	0.483	0.046	0.	0.	0.	0.	0.	0.	0.	89.38	
SAMPLE 22402 35	36.00	3.84	4.15	2.46	0.07	< 0.030	0.053	0.129	0.008	0.	0.	0.	0.	0.	0.	0.	96.74	

RELATIVE STANDARD DEVIATION %

INITIAL FLAME = 0.853 2.233 1.163 0.679 1.029 1.552 1.339 2.871 2.958

FINAL FLAME = 0.468 1.245 0.629 0.622 0.931 1.252 1.184 3.435 3.473

ABOUT LINE = 0.315 0.985 0.327 0.352 0.334 0.411 0.430 4.591 0.979

LIMIT DETECTION= 0.130 0.060 0.032 0.018 0.024 0.030 0.013 0.054 0.007

POLYNOMIAL ORDER

DRIFT = 8 7 9 8 8 2 6 3 1

CALIBRATION = 4 3 4 5 5 3 3 2 2

RELATIVE STANDARD DEVIATION ON STANDARD SY-2

BETWEEN SAMPLES=<0.685 1.416 0.697 <0.832 <1.423 <0.881 <0.983 <***** <2.069

MEASUREMENT

OBTAINED = 0.337 0.621 0.191 0.409 0.699 0.433 0.483 7.347 1.017

EXPECTED = 0.507 0.832 0.598 0.453 0.581 1.068 0.951 ***** 2.279

MEAN ON SY-2 = 59.96 12.08 6.21 2.65 7.95 4.282 4.481 0.157 0.326

ACCEPTED MEAN = 59.76 12.08 6.25 2.68 7.96 4.33 4.49 0.15 0.32

CaO

Na₂O + K₂O

SUMMARY OF STANDARD DEVIATION OF RESULTS

SAMPLE REF SIO2 AL203 FE203 MGO CAO NA20 K20 TIO2 MNO +H2O -H2O CO2 P205 S FEO FE203 TOTAL

SAMPLE 22369 1 0.241 0.141 0.080 0.032 0.074 0.024 0.007 0.049 0.005

SAMPLE 22370 2 0.239 0.

SAMPLE	22374	4	0.247	0.133	0.075	0.033	0.068	0.034	0.007	0.046	0.005	0.305
SAMPLE	22372	5	0.266	0.128	0.067	0.030	0.058	0.040	0.007	0.045	0.005	0.315
SAMPLE	22373	6	0.242	0.141	0.080	0.037	0.065	0.030	0.007	0.048	0.005	0.306
SAMPLE BLANK		7	0.066	0.030	0.016	0.006	0.007	0.015	0.007	0.025	0.003	0.080
SAMPLE	22375	8	0.225	0.130	0.075	0.043	0.040	0.019	0.011	0.044	0.005	0.281
SAMPLE	22376	9	0.317	0.054	0.133	0.029	0.010	0.015	0.007	0.026	0.004	0.351
SAMPLE SY-2	10	0.328	0.108	0.040	0.014	0.051	0.047	0.044	0.025	0.008	0.358	
SAMPLE	22377	11	0.213	0.124	0.068	0.035	0.052	0.015	0.014	0.045	0.005	0.268
SAMPLE	22378	12	0.221	0.117	0.070	0.032	0.060	0.022	0.009	0.044	0.005	0.274
SAMPLE	22379	13	0.227	0.120	0.074	0.036	0.053	0.023	0.008	0.043	0.006	0.279
SAMPLE	22381	14	0.240	0.136	0.077	0.033	0.076	0.029	0.007	0.046	0.005	0.304
SAMPLE	22382	15	0.228	0.124	0.069	0.028	0.064	0.031	0.007	0.043	0.005	0.282
SAMPLE	22383	16	0.254	0.131	0.078	0.032	0.061	0.042	0.007	0.046	0.005	0.310
SAMPLE	22384	17	0.206	0.129	0.070	0.031	0.049	0.015	0.016	0.043	0.004	0.264
SAMPLE BLANK	18	0.066	0.030	0.016	0.006	0.007	0.015	0.007	0.025	0.003	0.080	
SAMPLE	22385	19	0.297	0.043	0.033	0.307	0.010	0.021	0.010	0.036	0.005	0.432
SAMPLE	22386	20	0.224	0.121	0.070	0.031	0.045	0.015	0.012	0.044	0.005	0.274
SAMPLE	22387	21	0.250	0.126	0.072	0.026	0.068	0.033	0.007	0.044	0.005	0.303
SAMPLE SY-2	22	0.338	0.109	0.040	0.014	0.051	0.048	0.046	0.024	0.007	0.368	
SAMPLE	22389	23	0.238	0.141	0.082	0.037	0.062	0.033	0.007	0.049	0.005	0.303
SAMPLE	22390	24	0.237	0.137	0.063	0.035	0.047	0.036	0.008	0.047	0.005	0.298
SAMPLE	22391	25	0.234	0.128	0.067	0.033	0.032	0.016	0.016	0.047	0.005	0.284
SAMPLE	22393	26	0.206	0.134	0.063	0.034	0.044	0.032	0.008	0.048	0.005	0.272
SAMPLE	22394	27	0.213	0.120	0.060	0.027	0.048	0.016	0.020	0.044	0.005	0.263
SAMPLE	22395	28	0.211	0.117	0.067	0.031	0.030	0.015	0.011	0.044	0.005	0.259
SAMPLE	22396	29	0.236	0.144	0.078	0.033	0.062	0.030	0.008	0.049	0.005	0.301
SAMPLE BLANK	30	0.066	0.030	0.016	0.006	0.007	0.015	0.007	0.025	0.003	0.080	
SAMPLE	22397	31	0.235	0.141	0.086	0.036	0.059	0.037	0.007	0.048	0.005	0.302
SAMPLE	22398	32	0.232	0.139	0.075	0.033	0.069	0.026	0.008	0.049	0.005	0.296
SAMPLE	22399	33	0.239	0.134	0.074	0.031	0.076	0.022	0.007	0.048	0.005	0.300
SAMPLE SY-2	34	0.332	0.106	0.040	0.014	0.051	0.047	0.044	0.026	0.008	0.362	
SAMPLE	22401	35	0.427	0.082	0.032	0.015	0.008	0.015	0.007	0.027	0.004	0.437
SAMPLE	22402	36	0.642	0.043	0.029	0.013	0.007	0.015	0.007	0.025	0.003	0.645

end
-DONE

*BYE

**cost: \$ 6.85 to date: \$ 2018.93= 2%
**on at 15.150 - off at 15.897 on 01/09/80

MOORE BUSINESS FORMS • 7

CP DISCONNECTS

NPS DISCONNECTS]

SAMPLE NUMBER	REF NO	SUMMARY OF RESULTS IN % CORRECTED FOR BLANK AND BIAS								SET 75	ANALYST CHAUDHRY								
		SIO2	AL203	FE203	MGO	CAO	VA20	K20	TIO2		MNO	H120	H20	C02	P205	S	FEO	FE203	TOTAL
CC1 1245	1	47.53	14.90	9.80	6.34	5.86	3.760	0.152	1.466	0.172	0.	0.	0.	0.	0.	0.	0.	90.97	
• 129	2	46.10	15.37	9.40	6.35	7.73	2.407	1.014	1.482	0.151	0.	0.	0.	0.	0.	0.	0.	90.01	
• 67	3	43.43	15.52	16.18	9.75	3.72	1.180	0.050	1.961	0.203	0.	0.	0.	0.	0.	0.	0.	91.90	
• 112	4	74.51	7.41	6.71	0.29	0.21	0.042	0.942	0.338	0.009	0.	0.	0.	0.	0.	0.	0.	90.46	
• 53.3	5	75.72	9.76	3.29	1.15	1.80	0.184	1.988	0.636	0.271	0.	0.	0.	0.	0.	0.	0.	94.80	
• 63.7	6	46.04	13.11	10.10	6.73	8.81	0.080	1.477	1.578	0.161	0.	0.	0.	0.	0.	0.	0.	88.09	
SAMPLE BLANK	7	<0.09	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	0.25	
SAMPLE 22409	8	48.12	15.53	10.33	6.67	8.73	3.510	0.356	1.653	0.178	0.	0.	0.	0.	0.	0.	0.	95.15	
• 99	9	49.43	16.46	9.66	6.27	10.78	2.779	0.122	1.595	0.170	0.	0.	0.	0.	0.	0.	0.	97.27	
SAMPLE SY-2	10	60.80	12.22	5.14	2.63	7.99	4.257	4.642	0.140	0.317	0.46	0.23	0.46	0.44	0.01	3.54	2.20	100.23	
CC22 15.5	11	84.82	4.59	5.77	1.02	0.10	0.052	0.561	0.160	0.160	0.	0.	0.	0.	0.	0.	0.	97.28	
• 38.7	12	49.71	15.21	10.65	5.27	6.45	4.920	0.021	1.829	0.158	0.	0.	0.	0.	0.	0.	0.	94.22	
• 54.5	13	49.74	15.75	9.95	6.27	10.07	3.232	0.058	1.608	0.171	0.	0.	0.	0.	0.	0.	0.	96.84	
• 174.5	14	47.95	16.07	10.36	6.71	8.72	3.482	0.353	1.639	0.178	0.	0.	0.	0.	0.	0.	0.	94.81	
CC12 29.5	15	47.52	16.12	11.70	6.94	8.75	2.606	0.220	1.980	0.181	0.	0.	0.	0.	0.	0.	0.	96.11	
• 39.5	16	47.20	15.10	11.08	6.58	11.31	2.518	0.207	1.813	0.196	0.	0.	0.	0.	0.	0.	0.	96.00	
• 60.0	17	47.63	15.50	11.09	6.83	10.49	3.240	0.183	1.813	0.193	0.	0.	0.	0.	0.	0.	0.	96.98	
SAMPLE BLANK	18	<0.09	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	<0.02	0.25	
SAMPLE 22412	19	48.24	15.22	11.25	6.77	10.88	2.630	0.196	1.804	0.180	0.	0.	0.	0.	0.	0.	0.	97.15	
• 100	20	47.89	15.42	11.32	6.84	10.44	2.459	0.134	1.834	0.186	0.	0.	0.	0.	0.	0.	0.	96.52	
CC12 120	21	47.14	16.12	11.07	7.03	11.42	2.860	0.034	1.787	0.189	0.	0.	0.	0.	0.	0.	0.	97.65	
SAMPLE SY-2	22	60.35	12.35	6.13	2.68	7.90	4.289	4.451	0.147	0.318	0.46	0.23	0.46	0.44	0.01	3.54	2.20	99.83	
• 140	23	47.59	15.02	9.91	5.18	13.47	1.679	0.056	1.719	0.169	0.	0.	0.	0.	0.	0.	0.	94.78	
• 160	24	48.68	15.47	10.60	6.54	10.96	2.168	0.222	1.851	0.182	0.	0.	0.	0.	0.	0.	0.	96.68	
• 178	25	47.14	15.06	11.11	7.06	10.30	3.172	0.108	1.847	0.188	0.	0.	0.	0.	0.	0.	0.	95.99	
• 180	26	47.71	15.35	11.18	6.65	10.83	2.631	0.193	1.876	0.176	0.	0.	0.	0.	0.	0.	0.	96.32	
• 194.5	27	54.27	3.27	3.15	0.24	0.00?	0.033	0.064	0.012	0.490	0.021	0.	0.	0.	0.	0.	0.	0.	96.55
• 200.4	28	29.57	0.96	21.15	19.75	-	0.12	<0.007	<0.012	<0.052	0.039	0.	0.	0.	0.	0.	0.	0.	71.67
• 188.4	29	80.97	7.43	3.91	1.74	0.72	0.019	1.241	0.368	0.048	0.	0.	0.	0.	0.	0.	0.	96.44	
• 204.0	30	68.14	12.17	6.37	2.95	0.14	0.043	3.054	0.644	0.065	0.	0.	0.	0.	0.	0.	0.	93.57	
• 268.5	31	85.25	5.07	2.01	1.19	0.60	0.019	1.299	0.196	0.195	0.	0.	0.	0.	0.	0.	0.	95.72	
• 240.0	32	46.74	15.45	10.41	6.78	9.44	2.594	1.040	1.637	0.191	0.	0.	0.	0.	0.	0.	0.	94.27	
• 250.0	33	49.12	15.50	11.08	5.99	7.80	3.284	0.293	1.923	0.304	0.	0.	0.	0.	0.	0.	0.	95.38	
• 281.0	34	58.82	12.28	6.13	2.61	7.90	4.252	4.458	0.145	0.321	0.46	0.23	0.46	0.44	0.01	3.54	2.19	99.17	
• 350.0	35	56.74	14.49	8.02	4.28	7.31	3.651	0.232	1.494	0.172	0.	0.	0.	0.	0.	0.	0.	97.29	
SAMPLE 22433	36	49.03	15.78	11.04	5.94	7.88	3.253	0.293	1.902	0.305	0.	0.	0.	0.	0.	0.	0.	95.42	

RELATIVE STANDARD DEVIATION %
INITIAL FLAME = 1.029 0.702 0.783 0.974 0.779 1.110 1.447 2.949 2.906
FINAL FLAME = 0.467 0.601 0.505 0.862 0.851 0.833 0.974 4.021 3.091
ABOUT LINE = 0.220 1.666 0.350 0.818 0.601 0.587 0.315 **** 0.398
LIMIT DETECTION= 0.093 0.022 0.020 0.017 0.024 0.007 0.012 0.052 0.007
POLYNOMIAL ORDER
DRIFT = 8 6 7 10 8 6 7 9 6
CALIBRATION = 4 5 4 5 4 3 4 2 3
RELATIVE STANDARD DEVIATION ON STANDARD SY-2
BETWEEN SAMPLES= 0.745 <0.731 <0.648 <1.600 <0.697 <1.090 2.348 <5.921 <1.085
MEASUREMENT
OBTAINED = 0.153 0.359 0.318 0.786 0.343 0.536 0.459 2.910 0.533
EXPECTED = 0.398 0.415 0.488 0.585 0.556 0.648 0.803 **** 2.136
MEAN ON SY-2 = 60.35 12.28 6.13 2.64 7.89 4.266 4.517 0.144 0.319
ACCEPTED MEAN = 59.76 12.08 6.26 2.68 7.96 4.33 4.49 0.15 0.32

SAMPLE NUMBER	REF NO	SUMMARY OF STANDARD DEVIATION OF RESULTS										SET 75	ANALYST CHAUDHRY							
SIO2	AL203	FE203	MGO	CAO	VA20	K20	TIO2	MNO	H120	H20	C02	P205	S	FEO	FE203	TOTAL				

<tbl_r

SAMPLE	22406	4	0.347	0.050	0.035	0.007	0.012	0.003	0.010	0.028	0.003	0.354
SAMPLE	22407	5	0.358	0.064	0.019	0.011	0.016	0.003	0.016	0.030	0.006	0.367
SAMPLE	22408	6	0.192	0.080	0.061	0.047	0.053	0.003	0.013	0.052	0.005	0.235
SAMPLE	BLANK	7	0.047	0.010	0.010	0.005	0.011	0.003	0.005	0.024	0.003	0.057
SAMPLE	22409	8	0.204	0.103	0.064	0.047	0.052	0.023	0.007	0.054	0.005	0.255
SAMPLE	22410	9	0.210	0.116	0.058	0.043	0.066	0.018	0.006	0.053	0.005	0.265
SAMPLE	SY-2	10	0.271	0.075	0.033	0.020	0.047	0.030	0.042	0.026	0.007	0.294
SAMPLE	22411	11	0.408	0.032	0.031	0.011	0.012	0.003	0.008	0.027	0.005	0.412
SAMPLE	22412	12	0.213	0.101	0.068	0.036	0.039	0.038	0.006	0.059	0.004	0.260
SAMPLE	22413	13	0.211	0.104	0.060	0.043	0.061	0.021	0.007	0.053	0.005	0.261
SAMPLE	22414	14	0.201	0.109	0.065	0.047	0.052	0.023	0.007	0.053	0.005	0.255
SAMPLE	22415	15	0.200	0.108	0.077	0.049	0.052	0.017	0.007	0.063	0.005	0.259
SAMPLE	22416	16	0.200	0.100	0.074	0.046	0.070	0.016	0.006	0.058	0.005	0.258
SAMPLE	22417	17	0.202	0.104	0.073	0.048	0.064	0.021	0.006	0.058	0.005	0.259
SAMPLE	BLANK	18	0.047	0.010	0.010	0.006	0.011	0.003	0.006	0.024	0.003	0.057
SAMPLE	22418	19	0.205	0.102	0.075	0.047	0.067	0.017	0.006	0.061	0.005	0.263
SAMPLE	22419	20	0.202	0.102	0.075	0.048	0.064	0.016	0.006	0.059	0.005	0.259
SAMPLE	22420	21	0.198	0.108	0.071	0.049	0.070	0.019	0.007	0.058	0.005	0.259
SAMPLE	SY-2	22	0.270	0.075	0.033	0.020	0.047	0.030	0.040	0.026	0.007	0.292
SAMPLE	22421	23	0.201	0.097	0.060	0.035	0.085	0.011	0.007	0.056	0.005	0.255
SAMPLE	22422	24	0.207	0.103	0.067	0.046	0.067	0.014	0.007	0.059	0.005	0.262
SAMPLE	22423	25	0.199	0.098	0.073	0.050	0.063	0.021	0.006	0.059	0.005	0.255
SAMPLE	22424	26	0.203	0.106	0.075	0.046	0.066	0.017	0.006	0.059	0.005	0.262
SAMPLE	22425	27	0.227	0.031	0.021	5.620	0.014	0.004	0.008	0.033	0.004	5.624
SAMPLE	22426	28	0.125	0.016	0.181	0.195	0.013	0.003	0.007	0.029	0.004	0.297
SAMPLE	22427	29	0.384	0.050	0.022	0.015	0.013	0.003	0.012	0.028	0.003	0.389
SAMPLE	22428	30	0.310	0.075	0.034	0.022	0.012	0.003	0.025	0.032	0.004	0.325
SAMPLE	22429	31	0.409	0.039	0.015	0.012	0.012	0.003	0.012	0.027	0.005	0.412
SAMPLE	22430	32	0.193	0.094	0.061	0.046	0.056	0.017	0.011	0.055	0.005	0.242
SAMPLE	22431	33	0.205	0.096	0.068	0.040	0.047	0.021	0.007	0.062	0.007	0.253
SAMPLE	SY-2	34	0.259	0.078	0.033	0.020	0.047	0.029	0.039	0.029	0.007	0.283
SAMPLE	22432	35	0.243	0.088	0.050	0.029	0.044	0.024	0.007	0.051	0.005	0.274
SAMPLE	22433	36	0.204	0.098	0.067	0.040	0.047	0.021	0.007	0.062	0.007	0.253

end of file - request executed 96 times

-DONE
*BYE
**cost: \$ 2.14 to date: \$ 2107.74= 2%
**on at 8.122 - off at 8.342 on 01/11/80

CD DISCONNECTS

NPS DISCONNECTS/c)